



# PERRY JOHNSON LABORATORY ACCREDITATION, INC.

## *Certificate of Accreditation*

*Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:*

### ***DRL & Asociados, S.R.L.***

***Plaza ABBA, Avenida de Los Proceres No. 20, Urbanización Las Avenidas  
Santo Domingo DN, República Dominicana. C.P. 10601***

*(Hereinafter called the Organization) and hereby declares that Organization is accredited  
in accordance with the recognized International Standard:*

### **ISO/IEC 17025:2017**

This accreditation demonstrates technical competence for a defined scope and the  
operation of a laboratory quality management system  
(as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

### ***Mechanical Testing*** *(As detailed in the supplement)*

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen  
President

*Initial Accreditation Date:*

June 21, 2023

*Issue Date:*

June 21, 2023

*Expiration Date:*

September 30, 2025

*Accreditation No.:*

109973

*Certificate No.:*

L23-484

Perry Johnson Laboratory  
Accreditation, Inc. (PJLA)  
755 W. Big Beaver, Suite 1325  
Troy, Michigan 48084

*The validity of this certificate is maintained through ongoing assessments based on a  
continuous accreditation cycle. The validity of this certificate should be  
confirmed through the PJLA website: [www.pjlabs.com](http://www.pjlabs.com)*



# Certificate of Accreditation: Supplement

## DRL & Asociados, S.R.L.

Plaza ABBA, Avenida de Los Proceres No. 20, Urbanización Las Avenidas  
Santo Domingo DN, Republica Dominicana. C.P. 10601  
Contact Name: Walter de la Rosa Phone: 809-216-3938

Accreditation is granted to the facility to perform the following testing:

FIELD OF TEST	ITEMS, MATERIALS OR PRODUCTS TESTED	SPECIFIC TESTS OR PROPERTIES MEASURED	SPECIFICATION, STANDARD METHOD OR TECHNIQUE USED	RANGE (WHERE APPROPRIATE) AND DETECTION LIMIT
Mechanical Testing <sup>O</sup>	Biosafety Cabinets Clean Rooms (Particle Counts, Airflow Volume and Master, Temperature, Humidity Differential Pressure, Lux Meter, Flow Visualization and Integrity)	Design Qualification Installation Qualification Operating Qualification. Performance Qualification	Customer Requirement ISO 14644 DPD Air Hood and Air Meter, Digital Thermohigrometer, Manometer Lux Meter Smoke Candles DPD	0.3 µm to 10 µm 0.36 m/s to 0.54 m/s 18 °C to 25 °C 30 % RH to 75 % RH Up to 10 Pa 100 Lux to 2 000 Lux Visual 0.003 %, 0.005 %
	Fume Hood (Airflow Master, Differential Pressure, Flow Visualization Lux Meter)		Customer Requirement ASHRAE D-86331 Air Meter, Manometer Smoke Candles Lux Meter	0.36 m/s to 0.54 m/s Up to 10 Pa Visual 100 Lux to 2 000 Lux
	Metal Detector, Ferrous, Non-Ferrous, Stainless Steel	Particle sizes	ASTM 7046 Customer Requirement	0.01 mm to 10 mm Detect/ Non-Detect

- The presence of a superscript O means that the laboratory performs testing of the indicated parameter onsite at customer locations. Example: Outside Micrometer<sup>O</sup> would mean that the laboratory performs this testing onsite at the customer's location.